

# Aluminum Test Block

**Date 01082020**

**Form # 1100**


**Part # 14755**

This is to certify that the Magnaflux Liquid Penetrant Comparator, Aluminum Test Block meets the following requirements:

- ASME Boiler & Pressure Vessel Code 2019 Edition, Section V Nondestructive Examination, Article 6, Para. T-653 and Mandatory Appendices, Appendix III, Paragraph III-630
- AMS 2644G, Para. 4.4.5.2.1

#### Additional Information

- 1) No annual recertification is required.
- 2) Block must be ultrasonically cleaned and checked with a black light after each use.
- 3) Blocks that cannot be completely cleaned should be replaced.



Mathew Plamoottil

Quality Assurance Manager