

# Centrifuge Tube

Date 11112022

Form # 1100

Part # 507923

All Volumetric standards used for calibration are traceable to the National Institute of Standards and Technology with the following accuracy.

0-.2 mL	+/- .01 mL
At 25 mL	+/- 1 mL
Above 25 mL	+/- 2 mL



Roger St. Claire  
Quality Control Manager